

## **EC** Certificate

## Directive 93/42/EEC Annex II, excluding Section 4 Full Quality Assurance System Medical Devices

Registration No.: HD 60137356 0001

Report No.: 17039584 008

Manufacturer: Shenzhen Viatom Technology

Co., Ltd.

4E, Building 3, Tingwei Industrial Park

No. 6 Liufang Road, Block 67 Xin'an Street, Baoan District

Shenzhen

518101 Guangdong

China

**Products:** - Vital Signs Monitors

- Pulse Oximeters

- Blood Pressure Monitors

Replaces Approval, Registration No.: HD 60123955 0001

**Expiry Date:** 2024-05-27

**Effective Date:** 2019-07-17

**Date:** 2019-07-17

Fuxiu Sheng

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg
TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC
concerning medical devices with the identification number 0197.

Certification Department



TÜV Rheinland LGA Products GmbH • 51105 Köln

Shenzhen Viatom Technology Co., Ltd. 4E, Building 3, Tingwei Industrial Park, No.6 Liufang Road, Block 67, Xin'an Street, Baoan District, Shenzhen, 518101, Guangdong, P.R. China

Contact

Tel. +49 911 655-5225 Mail: medicalproducts@de.tuv.com

Date January 22, 2024

**Notified Body Confirmation Letter** 

: 10924064 Reference.

To whom it may concern,

Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices.

This letter confirms that TÜV Rheinland LGA Products GmbH, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number 0197 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

Shenzhen Viatom Technology Co., Ltd. 4E, Building 3, Tingwei Industrial Park, No.6 Liufang Road, Block 67, Xin'an Street, Baoan District, Shenzhen, 518101, Guangdong, P.R. China SRN Number (if available): CN-MF-000012182

The devices covered by the formal application and the written agreement mentioned above are identified in the tables below. Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which the NB is also responsible for appropriate surveillance under the applicable Directive. Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but the NB has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after May 26, 2021 but before March 20, 2023 without having been withdrawn, this letter also confirms that the manufacturer either signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry; or provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively, by March 20, 2023 for the relevant devices.

TÜV Rheinland LGA Products GmbH

Am Grauen Stein 51105 Köln Germany

Headquarter

Tillystraße 2 90431 Nuremberg

Phone. +49 911 655 5225 +49 911 655 5226 service@de tuy com www.tuv.com/safety

**Board of Management** 

Dipl.-Ing. Thomas Weigand, Spokesman

Dipl.-Kfm. Dr. Jörg Schlösser

Nuremberg HRB 26013 VAT No.: DE 811835490

Chairman of the Supervisory Board

Dr.-Ing. Michael Fübi

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3c of MDR (as amended by (EU) 2023/607), are shown below:

- May 26, 2026 for Class III custom-made implantable devices
- December 31, 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- December 31, 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- December 31, 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body

Samuel Qin Certification body

Table 1: Devices covered by this letter and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

**Device name or Basic** MDR Device If the MDR device MDD/AIMDD **UDI-DI** (under MDR classification (as is a substitute Certificate application) proposed by the device. Reference(s) of manufacturer identification of the devices and verified at the corresponding under MDR the pre-MDD/AIMDD application, and application device the NB stage) Identification **Vital Signs Monitor** Class IIa Health Monitor Certificate # Model: Model: HD 60137356 Checkme Pro. Checkme Pro. 0001 Checkme Plus, Checkme Plus, Checkme Pod. Checkme Pod. NB #0197 Checkme Lite Checkme Lite Basic UDI-DI: 69344401CheckmePro9X **Blood Pressure Monitor** Class IIa N/A Certificate # Model: HD 60137356 BP1, BP1A, BP1S, 0001 BP1SA, BP1B, BP1C NB #0197 **Basic UDI-DI:** 69344401BP1SP8 **Blood Pressure Monitor** Class IIa N/A Certificate # Model: HD 60137356 BP2A 0001 **Basic UDI-DI:** NB #0197 69344401BP2AN7

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the preapplication stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Pulse Oximeter Model: Oxiband  Basic UDI-DI: 69344401Oxiband29	Class IIa	N/A	Certificate # HD 60137356 0001 NB #0197
Pulse Oximeter Model: PO2, PO4 Basic UDI-DI: 69344401PO2ZG	Class IIa	N/A	Certificate # HD 60137356 0001 NB #0197
Pulse Oximeter Model: PO6, PO6A, PO1, PO1B Basic UDI-DI: 69344401PO6ZQ	Class IIa	N/A	Certificate # HD 60137356 0001 NB #0197
Pulse Oximeter Model: VTM01, VTM01A, VTM01B  Basic UDI-DI: 69344401VTM01S7	Class IIa	N/A	Certificate # HD 60137356 0001 NB #0197
Pulse Oximeter Model: PO6B, PO6C Basic UDI-DI: 69344401PO6BRJ	Class IIa	N/A	Certificate # HD 60137356 0001 NB #0197
Pulse Oximeter Model: PF-10AW, PF-10AW1, PF-10A, PF-10A1, PF-10BW, PF-10BW1, PF-10B, PF-10B1  Basic UDI-DI: 69344401PF-10AW9D	Class IIa	N/A	Certificate # HD 60137356 0001 NB #0197
Pulse Oximeter Model: PF-20AW, PF-20AW1, PF-20B, PF-20B1  Basic UDI-DI: 69344401PF-20AW9L	Class IIa	N/A	Certificate # HD 60137356 0001 NB #0197

UDI-DI (under MDR application) classification (as proposed by the manufacturer and verified at the pre-	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
---	--	--

Table 2: Devices covered by this letter and for which the NB is <u>NOT</u> responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Device name or	MDR Device	If the MDR device	MDD/AIMDD
Basic UDI-DI	classification (as	is a substitute	Certificate
(under MDR	proposed by the	device,	Reference(s) of the
application)	manufacturer and	identification of the	devices under MDR
	verified at the pre-	corresponding	application, and the
	application stage)	MDD/AIMDD device	NB Identification
N/A	N/A	N/A	N/A

**Confirmation Letter Revision History** 

Date	NB internal reference traceable to each version of the letter	Action
2024-01-22	10924064	Initial issue